

DECLARATION OF CONFORMITY

We

UEBE Medical GmbH
Bgm.-Kuhn-Str. 22
97900 Kulsheim / Germany

Declare under our sole responsibility that the medical software application (App)

cyclotest[®] mySense App REF 064001 App Version 2.XX and onwards

meets all the provisions of the directive 93/42/EEC which apply to it

Classification according Annex IX of MDD 93/42/EEC			
Class			Applied Rule According Annex IX
I <input type="checkbox"/>	IIa <input type="checkbox"/>	IIb <input checked="" type="checkbox"/>	Rule 14

Rule 14 because cyclotest mySense App is a fertility analysis application to determine the fertile and infertile days in cycles of women.

Conformity assessment procedure according to Annex II without 4 of the directive 93/42/EEC

Notified Body:

TÜV SÜD Product Service GmbH *Notified Body CE 0123*
 Ridlerstraße 65
 80339 München

Certificate No. G1 029547 0027 Rev. 01

Valid until: 26.05.2024

Kulsheim, 09.05.2022



Alexander Stiefenhofer
 Person Responsible For Regulatory Compliance